

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Collagen R Lösung 0.2 %, steril	Kat.Nr./Cat.No. : 47254
Collagen R solution 0.2 %, sterile	Lot/Contr.No. : 150593

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Molekulargewicht molecular weight		ca. 300 000	
Aussehen appearance		klare, viskose Lösung clear, viscous solution	entspricht corresponds
Gehalt (%) assay	UV	0.18 – 0.22	0.22
pH		3.2 – 3.8	3.6
Steriltest sterile test		kein Wachstum no growth	entspricht corresponds
Gelbildung gel formation		positiv positive	entspricht corresponds
Ursprung origin		Rattenschwanz rat tail	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			04/ 2017
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 01.06.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.