

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Mycorase-Lösung (50X) für Mykoplasmen-Entfernung	Kat.Nr./Cat.No. : 47987
Mycorase solution (50x) for mycoplasma removal	Lot/Contr.No. : 201164

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	gelbliche Flüssigkeit yellowish liquid	entspricht corresponds
pH		9.34
Osmolalität (mOsmol/ kg) osmolality		272
Mindesthaltbarkeit minimum shelf life		28.01.2022
Lagerung (°C) storage		-15 bis/ to -25

SERVA Electrophoresis GmbH

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 03.11.2020

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.