

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Mycorase-Lösung (50X) für Mykoplasmen-Entfernung</b>	<b>Kat.Nr./Cat.No.:</b>	<b>47987</b>
Mycorase solution (50x) for mycoplasma removal	<b>Lot/Contr.No.:</b>	<b>230807</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	gelbliche Flüssigkeit yellowish liquid	entspricht corresponds
<b>pH</b>		9.6
<b>Osmolalität (mOsmol/ kg)</b> osmolality		277
<b>Mindesthaltbarkeit</b> minimum shelf life		04.04.2025
<b>Lagerung (°C)</b> storage		-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 27.07.2023**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.