

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

M9 Minimal salts, 5x Pulvermedium
powder

Kat.Nr./Cat.No. : 48505
Lot/Contr.No. : P230293

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
Löslichkeit solubility	52.5 g/l in H ₂ O	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	52.5 g/l in H ₂ O	6.6 – 6.8	6.8
Osmolarität (mOsmol/l) osmolarity	52.5 g/l in H ₂ O	700 – 800	794
Zusammensetzung composition			
Na ₂ HPO ₄		30.0 g/l	entspricht corresponds
KH ₂ PO ₄		15.0 g/l	entspricht corresponds
NH ₄ Cl		5.0 g/l	entspricht corresponds
NaCl		2.5 g/l	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			25.07.2026
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 02.08.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.