

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Trypton aus Casein, pankreatisch	Kat.Nr./Cat.No. : 48647
Tryptone from casein, pancreatic	Lot/Contr.No. : 250326

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		cremefarbenes Pulver cream colored powder	entspricht corresponds
pH	2% in H ₂ O	6.5 – 7.5	7.1
Trocknungsverlust (%) loss on drying		max. 6.0	2.9
Gesamtstickstoff (TN) (%) total nitrogen		min. 10.00	12.89
Aminostickstoff (AN) (%) amino nitrogen		min. 3.90	5.42
AN/TN x 100		report	42.05
Asche (%) ash		max. 15.0	6.2
Mikrobiologische Analyse microbiological analysis	standard plate count	< 5 000 CFU/g	entspricht/ corresponds
	Salmonella	negative	entspricht/ corresponds
	Coliforms	negative	entspricht/ corresponds
	Yeasts and moulds	< 100 CFU/g	entspricht/ corresponds
Mindesthaltbarkeit minimum shelf life			04.07.2028
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 08.04.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.